
Midmark M11 - v1.0.5
Total Cycles: 1129

Name: RITTER2
Sterilizer ID: V2313228
05/27/2022 10:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.0 F 0.0
2:00 197.0 F 4.3
4:00 210.6 F 10.4
6:00 226.5 F 6.5
8:00 238.8 F 13.1
10:00 240.7 F 11.7
12:00 241.0 F 11.3
14:00 255.2 F 18.5
16:00 267.2 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 271.5 F 29.3
2:00 272.8 F 29.6
3:00 271.8 F 29.1
4:00 271.1 F 28.8

Min 271.1 F 28.7
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 230.9 F 0.0
10:00 218.3 F 0.0
15:00 214.2 F 0.0
20:00 215.2 F 0.0
25:00 217.4 F 0.0
30:00 218.4 F 0.0
Drying Complete

FILLING: 2:28
HEATING: 17:20
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:56:57
CYCLE COMPLETE

05/27/2022 11:39 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____