
Midmark M11 - v1.0.5
Total Cycles: 1412

Name: RITTER2
Sterilizer ID: V2313228
08/25/2022 07:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.5 F 0.0
2:00 74.7 F 0.0
4:00 81.5 F 0.0
6:00 101.4 F 0.1
8:00 144.8 F 1.7
10:00 170.0 F 3.5
12:00 196.2 F 5.9
14:00 214.1 F 9.4
16:00 225.2 F 13.2
18:00 226.0 F 8.1
20:00 237.8 F 12.5
22:00 236.0 F 9.6
24:00 238.7 F 9.6
26:00 247.4 F 14.3
28:00 259.1 F 20.8
30:00 269.7 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.3 F 29.2
2:00 271.5 F 29.2
3:00 272.1 F 29.0
4:00 271.8 F 29.3

Min 271.3 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.7 F 0.0
5:00 211.7 F 0.0
10:00 226.9 F 0.0
15:00 221.3 F 0.0
20:00 228.4 F 0.0
25:00 229.2 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:56
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:08:48
CYCLE COMPLETE

08/25/2022 08:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____