
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000376
13:13 21/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:07
STERILIZING 3:13
136.0 C 220kPa 3:13
136.2 C 226kPa 3:43
136.3 C 231kPa 4:13
136.1 C 229kPa 4:43
136.5 C 230kPa 5:13
136.4 C 229kPa 5:43
136.4 C 231kPa 6:13
136.3 C 231kPa 6:43
136.3 C 231kPa 6:43
VENTING 6:43
AIR DRYING 7:35
CYCLE COMPLETE 7:40

Digital Signature #
1F32D90FD86ED6F5
