
Midmark M11 - v1.0.5
Total Cycles: 3220

Name: RITTER2
Sterilizer ID: V2313228
12/14/2023 05:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.5 F 0.0
2:00 156.3 F 2.9
4:00 182.1 F 6.9
6:00 204.8 F 12.1
8:00 220.6 F 6.3
10:00 233.7 F 11.7
12:00 234.6 F 9.3
14:00 238.2 F 9.6
16:00 248.8 F 15.1
18:00 261.5 F 22.1
20:00 272.3 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.7 F 29.6
2:00 271.7 F 28.8
3:00 271.7 F 29.3
4:00 272.7 F 29.5

Min 271.1 F 28.6
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.7 F 0.0
5:00 193.3 F 0.0
10:00 196.0 F 0.0
15:00 193.2 F 0.0
20:00 188.8 F 0.0
25:00 186.8 F 0.0
30:00 184.9 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 20:31
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:59:19
CYCLE COMPLETE

12/14/2023 06:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____