
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 001463
12:23 09/30/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:12
PRESSURIZING 6:58
STERILIZING 8:37
136.4 C 222kPa 8:37
136.3 C 222kPa 9:07
135.8 C 219kPa 9:37
136.2 C 218kPa 10:07
136.3 C 221kPa 10:37
136.3 C 222kPa 11:07
136.0 C 221kPa 11:37
136.5 C 223kPa 12:07
VENTING 12:07
AIR DRYING 13:11
CYCLE COMPLETE 15:41

Digital Signature #
381875BD2B7EAB14
