

-----  
Midmark M11 - v1.0.5  
Total Cycles: 971

Name: RITTER2  
Sterilizer ID: V2313228  
04/11/2022 07:17 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.3 F 0.0  
2:00 74.4 F 0.0  
4:00 94.9 F 0.0  
6:00 130.2 F 0.0  
8:00 165.9 F 0.1  
10:00 170.0 F 0.4  
12:00 203.4 F 1.1  
14:00 213.8 F 4.6  
16:00 223.2 F 7.9  
18:00 234.1 F 11.6  
20:00 233.7 F 8.1  
22:00 243.7 F 13.4  
24:00 242.7 F 12.0  
26:00 241.7 F 11.3  
28:00 254.4 F 17.8  
30:00 265.4 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.6  
1:00 272.4 F 29.3  
2:00 271.3 F 28.5  
3:00 272.4 F 29.4  
4:00 271.9 F 28.7

Min 271.2 F 28.5  
Max 272.5 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.9 F 0.0  
5:00 195.7 F 0.0  
10:00 193.5 F 0.0  
15:00 187.4 F 0.0  
20:00 192.3 F 0.0  
25:00 204.3 F 0.0  
30:00 218.6 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 31:49  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:09:55  
CYCLE COMPLETE

04/11/2022 08:26 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_