

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001479  
14:49 25/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:33  
PRESSURIZING 2:36  
STERILIZING 3:45  
136.0 C 220kPa 3:45  
136.2 C 226kPa 4:15  
136.2 C 223kPa 4:45  
136.3 C 225kPa 5:15  
136.4 C 225kPa 5:45  
136.4 C 226kPa 6:15  
136.4 C 226kPa 6:45  
136.3 C 225kPa 7:15  
136.3 C 225kPa 7:15  
VENTING 7:15  
AIR DRYING 8:10  
CYCLE COMPLETE 9:40  
-----

Digital Signature #  
52792C30D6D6CAD9  
-----