
Midmark M11 - v1.0.5
Total Cycles: 3564

Name: RITTER2
Sterilizer ID: V2313228
04/26/2024 11:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.9 F 0.0
2:00 146.3 F 1.8
4:00 156.9 F 4.6
6:00 185.2 F 8.1
8:00 224.3 F 12.0
10:00 214.8 F 4.8
12:00 224.5 F 8.5
14:00 233.6 F 12.6
16:00 235.7 F 9.2
18:00 240.8 F 11.1
20:00 245.4 F 12.9
22:00 256.0 F 18.6
24:00 266.1 F 24.7
26:00 270.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 273.3 F 29.6
2:00 271.7 F 29.2
3:00 272.4 F 29.6
4:00 271.6 F 29.0

Min 271.6 F 29.0
Max 273.3 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.2 F 0.0
5:00 235.4 F 0.0
10:00 219.9 F 0.0
15:00 178.1 F 0.0
20:00 177.4 F 0.0
25:00 175.1 F 0.0
30:00 173.8 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 26:47
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:05:38
CYCLE COMPLETE

04/26/2024 12:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____