

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 006783  
9:10 27/12/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:03  
PRESSURIZING 3:05  
STERILIZING 4:03  
136.0 C 220kPa 4:03  
135.9 C 220kPa 4:33  
136.4 C 222kPa 5:03  
136.4 C 223kPa 5:33  
136.1 C 221kPa 6:03  
136.0 C 220kPa 6:33  
136.2 C 221kPa 7:03  
136.4 C 220kPa 7:33  
136.4 C 220kPa 7:33  
VENTING 7:33  
AIR DRYING 8:20  
CYCLE COMPLETE 38:20  
-----

Digital Signature #  
2F7F8EB73F8F044A  
-----