

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001306  
16:04 19/10/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:27  
PRESSURIZING 2:36  
STERILIZING 3:41  
136.0 C 220kPa 3:41  
136.2 C 220kPa 4:11  
136.1 C 222kPa 4:41  
136.3 C 224kPa 5:11  
136.4 C 224kPa 5:41  
136.3 C 223kPa 6:11  
136.3 C 224kPa 6:41  
136.4 C 224kPa 7:11  
136.4 C 224kPa 7:11  
VENTING 7:11  
AIR DRYING 8:03  
CYCLE COMPLETE 15:59  
-----

Digital Signature #  
214BE2179EE292E1  
-----