
Midmark M11 - v1.0.5
Total Cycles: 2723

Name: RITTER2
Sterilizer ID: V2313228
08/02/2023 08:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 70.8 F 0.0
4:00 75.6 F 0.0
6:00 98.1 F 0.0
8:00 153.9 F 0.0
10:00 182.7 F 0.0
12:00 202.3 F 0.1
14:00 213.7 F 3.3
16:00 222.5 F 6.1
18:00 230.6 F 9.3
20:00 239.5 F 13.1
22:00 237.2 F 9.9
24:00 238.7 F 9.6
26:00 246.1 F 12.8
28:00 245.2 F 12.9
30:00 257.1 F 19.3
32:00 268.0 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.7 F 29.0
2:00 271.9 F 28.8
3:00 272.4 F 28.8
4:00 272.0 F 28.9

Min 271.6 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.4 F 0.0
5:00 216.8 F 0.0
10:00 219.9 F 0.0
15:00 222.3 F 0.0
20:00 213.8 F 0.0
25:00 215.5 F 0.0
30:00 215.6 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 33:17
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:11:18
CYCLE COMPLETE

08/02/2023 09:44 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____