
Midmark M11 - v1.0.5
Total Cycles: 3758

Name: RITTER2
Sterilizer ID: V2313228
07/08/2024 09:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 67.7 F 0.0
2:00 71.2 F 0.0
4:00 112.3 F 0.0
6:00 115.5 F 0.5
8:00 140.9 F 3.1
10:00 170.0 F 5.9
12:00 192.2 F 8.8
14:00 204.4 F 12.8
16:00 214.6 F 5.7
18:00 226.0 F 9.5
20:00 237.8 F 12.9
22:00 235.7 F 10.1
24:00 238.8 F 9.2
26:00 245.9 F 13.6
28:00 256.9 F 19.6
30:00 267.0 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.7 F 29.7
2:00 271.3 F 28.9
3:00 273.0 F 29.8
4:00 271.7 F 29.0

Min 271.3 F 28.9
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.6 F 0.0
5:00 195.8 F 0.0
10:00 197.5 F 0.0
15:00 193.6 F 0.0
20:00 203.5 F 0.0
25:00 207.7 F 0.0
30:00 206.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:34
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:09:48
CYCLE COMPLETE

07/08/2024 10:28 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____