
Midmark M11 - v1.0.0
Total Cycles: 4835

Name: RITTER 1
Sterilizer ID: V2178198
08/20/2024 10:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.6 F 0.0
2:00 98.8 F 0.0
4:00 127.2 F 0.4
6:00 149.7 F 1.3
8:00 178.7 F 2.9
10:00 201.1 F 6.6
12:00 215.1 F 10.9
14:00 220.3 F 5.3
16:00 230.7 F 9.3
18:00 238.0 F 10.4
20:00 241.6 F 11.5
22:00 238.5 F 9.6
24:00 252.1 F 16.2
26:00 263.5 F 22.8
28:00 271.9 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.7 F 28.8
2:00 271.5 F 28.0
3:00 272.7 F 28.9
4:00 272.0 F 28.3

Min 271.5 F 28.0
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.9 F 0.0
5:00 198.5 F 0.0
10:00 193.2 F 0.0
15:00 211.3 F 0.0
20:00 216.6 F 0.0
25:00 222.7 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 28:10
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:06:23
CYCLE COMPLETE

08/20/2024 11:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____