
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000349
8:29 20/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:12
PRESSURIZING 2:22
STERILIZING 3:22
136.0 C 225kPa 3:22
136.2 C 232kPa 3:52
136.1 C 230kPa 4:22
136.2 C 230kPa 4:52
136.1 C 231kPa 5:22
136.4 C 232kPa 5:52
136.4 C 229kPa 6:22
136.3 C 232kPa 6:52
136.3 C 232kPa 6:52
VENTING 6:52
AIR DRYING 7:43
CYCLE COMPLETE 7:48

Digital Signature #
F851B39263718BDC
