
Midmark M11 - v1.0.0
Total Cycles: 4162

Name: RITTER 1
Sterilizer ID: V2178198
12/28/2023 09:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.7 F 0.0
2:00 76.3 F 0.3
4:00 88.3 F 1.5
6:00 118.8 F 3.7
8:00 156.5 F 6.4
10:00 178.1 F 9.7
12:00 202.1 F 13.5
14:00 211.5 F 4.3
16:00 221.2 F 7.6
18:00 230.4 F 11.2
20:00 228.7 F 6.5
22:00 240.2 F 11.3
24:00 235.4 F 8.5
26:00 248.5 F 14.4
28:00 259.3 F 20.3
30:00 269.2 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.0 F 28.5
2:00 272.6 F 28.7
3:00 272.5 F 29.1
4:00 272.0 F 28.9

Min 271.6 F 28.4
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.2 F 0.0
5:00 232.5 F 0.0
10:00 227.9 F 0.0
15:00 229.1 F 0.0
20:00 224.9 F 0.0
25:00 231.0 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:43
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:09:58
CYCLE COMPLETE

12/28/2023 10:36 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____