
Midmark M11 - v1.0.0
Total Cycles: 2164

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2022 01:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.8 F 0.0
2:00 174.5 F 3.7
4:00 196.4 F 8.3
6:00 212.7 F 13.3
8:00 223.9 F 7.0
10:00 236.0 F 12.0
12:00 234.4 F 8.8
14:00 239.8 F 10.2
16:00 246.0 F 13.4
18:00 258.2 F 19.9
20:00 268.9 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.2
2:00 271.8 F 28.5
3:00 271.9 F 29.0
4:00 272.6 F 29.1

Min 271.4 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.6 F 0.0
5:00 220.0 F 0.0
10:00 225.9 F 0.0
15:00 226.7 F 0.0
20:00 224.2 F 0.0
25:00 222.3 F 0.0
30:00 216.9 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 21:09
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

07/21/2022 02:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____