

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1369

Name: RITTER 1  
Sterilizer ID: V2178198  
11/15/2021 11:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.1 F 0.0  
2:00 173.3 F 3.2  
4:00 195.7 F 7.7  
6:00 218.4 F 13.3  
8:00 228.0 F 8.3  
10:00 238.6 F 11.6  
12:00 240.2 F 11.4  
14:00 238.5 F 9.9  
16:00 252.4 F 16.8  
18:00 264.3 F 23.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.6 F 29.0  
2:00 271.9 F 29.0  
3:00 272.3 F 28.8  
4:00 272.0 F 29.0  
  
Min 271.4 F 28.5  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
5:00 227.4 F 0.0  
10:00 221.4 F 0.0  
15:00 223.6 F 0.0  
20:00 221.0 F 0.0  
25:00 225.9 F 0.0  
30:00 223.1 F 0.0  
Drying Complete

FILLING: 1:51  
HEATING: 19:53  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:59:00  
CYCLE COMPLETE

11/15/2021 12:28 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02