
Midmark M11 - v1.0.5
Total Cycles: 2772

Name: RITTER2
Sterilizer ID: V2313228
08/15/2023 03:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.1 F 0.0
2:00 200.8 F 1.2
4:00 210.7 F 6.9
6:00 229.0 F 12.5
8:00 232.4 F 9.2
10:00 238.8 F 9.8
12:00 244.0 F 12.3
14:00 245.9 F 13.5
16:00 258.9 F 20.6
18:00 270.2 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.3 F 29.0
2:00 272.8 F 29.5
3:00 271.8 F 29.1
4:00 272.2 F 29.0

Min 271.6 F 28.7
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 219.6 F 0.0
10:00 225.9 F 0.0
15:00 227.1 F 0.0
20:00 226.2 F 0.0
25:00 226.0 F 0.0
30:00 222.4 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 18:47
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

08/15/2023 04:13 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____