
Midmark M11 - v1.0.5
Total Cycles: 1645

Name: RITTER2
Sterilizer ID: V2313228
10/24/2022 12:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.3 F 0.0
2:00 152.8 F 2.5
4:00 165.6 F 6.1
6:00 196.6 F 10.3
8:00 210.9 F 4.1
10:00 224.0 F 9.1
12:00 235.6 F 11.2
14:00 238.1 F 11.1
16:00 236.8 F 9.3
18:00 249.7 F 15.8
20:00 261.0 F 22.2
22:00 271.3 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 271.5 F 29.7
2:00 272.8 F 29.8
3:00 271.6 F 29.3
4:00 271.3 F 29.6

Min 271.1 F 29.1
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 206.2 F 0.0
10:00 218.6 F 0.0
15:00 211.8 F 0.0
20:00 208.2 F 0.0
25:00 206.0 F 0.0
30:00 207.6 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 22:34
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:01:09
CYCLE COMPLETE

10/24/2022 01:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____