

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000792  
10:16 11/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:09  
PRESSURIZING 3:11  
STERILIZING 4:16  
136.3 C 221kPa 4:16  
136.3 C 224kPa 4:46  
136.1 C 222kPa 5:16  
136.4 C 223kPa 5:46  
136.2 C 220kPa 6:16  
136.1 C 220kPa 6:46  
136.3 C 224kPa 7:16  
136.4 C 225kPa 7:46  
136.4 C 225kPa 7:46  
VENTING 7:46  
AIR DRYING 8:42  
CYCLE COMPLETE 8:44

-----  
Digital Signature #  
CE65A9BDD54A8EB4  
-----