

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2716

Name: RITTER2  
Sterilizer ID: V2313228  
08/01/2023 09:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.9 F 0.0  
2:00 180.7 F 0.0  
4:00 195.2 F 3.9  
6:00 213.7 F 8.9  
8:00 226.0 F 9.7  
10:00 236.2 F 11.0  
12:00 236.1 F 9.4  
14:00 237.7 F 9.1  
16:00 251.1 F 16.1  
18:00 263.5 F 23.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.5 F 29.3  
2:00 271.6 F 29.0  
3:00 272.7 F 29.4  
4:00 271.6 F 28.7  
  
Min 271.5 F 28.7  
Max 273.0 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.4 F 0.0  
5:00 205.3 F 0.0  
10:00 220.2 F 0.0  
15:00 219.4 F 0.0  
20:00 220.0 F 0.0  
25:00 218.3 F 0.0  
30:00 216.6 F 0.0  
  
Drying Complete

FILLING: 1:44  
HEATING: 19:60  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:54  
CYCLE COMPLETE

08/01/2023 10:49 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_