
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001184
8:55 17/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 1:57
STERILIZING 2:51
136.5 C 220kPa 2:51
136.0 C 221kPa 3:21
136.4 C 220kPa 3:51
136.1 C 221kPa 4:21
136.4 C 223kPa 4:51
136.4 C 224kPa 5:21
CYCLE ABORTED 5:33
STOP BUTTON PRESSED
NOT STERILE
END TIME 5:33

Digital Signature #
EC713DAB77DC584D
