
Midmark M11 - v1.0.0
Total Cycles: 2635

Name: RITTER 1
Sterilizer ID: V2178198
12/05/2022 09:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.1 F 0.0
2:00 76.8 F 0.0
4:00 89.1 F 0.0
6:00 116.3 F 0.4
8:00 150.0 F 1.9
10:00 174.6 F 4.7
12:00 194.3 F 7.8
14:00 209.9 F 11.9
16:00 218.1 F 5.0
18:00 227.2 F 9.4
20:00 237.7 F 11.0
22:00 239.5 F 10.7
24:00 235.4 F 8.3
26:00 249.6 F 15.0
28:00 260.8 F 21.4
30:00 271.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.4
1:00 271.6 F 28.3
2:00 271.6 F 28.5
3:00 271.6 F 28.7
4:00 271.7 F 28.7

Min 271.4 F 28.3
Max 272.3 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 220.9 F 0.0
5:00 205.2 F 0.0
10:00 229.1 F 0.0
15:00 223.7 F 0.0
20:00 220.7 F 0.0
25:00 220.4 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 30:26
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:21
CYCLE COMPLETE

12/05/2022 10:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____