
Midmark M11 - v1.0.0
Total Cycles: 3561

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2023 11:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.6 F 0.0
2:00 148.1 F 0.0
4:00 157.9 F 0.8
6:00 188.5 F 3.2
8:00 221.9 F 6.4
10:00 216.3 F 10.3
12:00 228.2 F 8.1
14:00 229.6 F 9.0
16:00 240.2 F 13.5
18:00 239.9 F 10.7
20:00 235.9 F 8.6
22:00 249.8 F 15.1
24:00 260.8 F 21.3
26:00 270.9 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.3
1:00 272.7 F 29.2
2:00 271.7 F 28.6
3:00 272.5 F 28.9
4:00 272.3 F 29.3

Min 271.0 F 28.2
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.5 F 0.0
5:00 234.2 F 0.0
10:00 195.6 F 0.0
15:00 187.3 F 0.0
20:00 184.7 F 0.0
25:00 182.0 F 0.0
30:00 181.3 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 26:48
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:05:11
CYCLE COMPLETE

07/21/2023 12:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____