

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4760

Name: RITTER 1  
Sterilizer ID: V2178198  
07/23/2024 01:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 84.9 F 0.0  
2:00 88.0 F 0.2  
4:00 117.0 F 1.5  
6:00 140.7 F 4.0  
8:00 172.6 F 6.6  
10:00 186.0 F 10.1  
12:00 208.5 F 9.1  
14:00 217.2 F 4.9  
16:00 227.5 F 8.3  
18:00 236.5 F 12.2  
20:00 233.6 F 8.1  
22:00 244.0 F 13.2  
24:00 242.0 F 11.3  
26:00 253.4 F 17.0  
28:00 263.8 F 23.1  
30:00 272.4 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.7 F 29.0  
2:00 271.9 F 28.2  
3:00 272.1 F 28.8  
4:00 272.6 F 28.8  
  
Min 271.5 F 28.1  
Max 272.9 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.0 F 0.0  
5:00 196.9 F 0.0  
10:00 206.8 F 0.0  
15:00 207.7 F 0.0  
20:00 213.0 F 0.0  
25:00 215.2 F 0.0  
30:00 215.5 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:16  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:08:50  
CYCLE COMPLETE

07/23/2024 02:50 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_