
Midmark M11 - v1.0.5
Total Cycles: 2329

Name: RITTER2
Sterilizer ID: V2313228
05/03/2023 09:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.6 F 0.0
2:00 135.9 F 1.5
4:00 146.4 F 4.5
6:00 176.0 F 8.7
8:00 200.1 F 13.5
10:00 217.8 F 6.9
12:00 228.8 F 12.1
14:00 231.6 F 8.9
16:00 238.0 F 10.3
18:00 245.0 F 13.6
20:00 257.4 F 20.3
22:00 268.2 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.5
1:00 272.1 F 30.0
2:00 271.2 F 29.5
3:00 272.1 F 30.1
4:00 271.4 F 29.4

Min 271.2 F 29.3
Max 272.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.5 F 0.0
5:00 207.5 F 0.0
10:00 213.6 F 0.0
15:00 216.7 F 0.0
20:00 219.1 F 0.0
25:00 221.4 F 0.0
30:00 222.4 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:11
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:01:27
CYCLE COMPLETE

05/03/2023 10:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____