
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000203
13:14 21/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:44
PRESSURIZING 3:49
STERILIZING 5:24
136.0 C 229kPa 5:24
136.3 C 232kPa 5:54
136.3 C 232kPa 6:24
136.4 C 233kPa 6:54
136.4 C 233kPa 7:24
136.4 C 233kPa 7:54
136.4 C 233kPa 8:24
136.4 C 233kPa 8:54
136.4 C 233kPa 8:54
VENTING 8:54
AIR DRYING 10:03
CYCLE COMPLETE 25:03

Digital Signature #
2AE6AADFD27F330D
