
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000178
6:46 17/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:23
PRESSURIZING 4:24
STERILIZING 5:41
136.0 C 220kPa 5:41
136.5 C 225kPa 6:11
136.4 C 225kPa 6:41
136.3 C 223kPa 7:11
136.4 C 224kPa 7:41
136.4 C 224kPa 8:11
136.4 C 223kPa 8:41
136.1 C 221kPa 9:11
136.1 C 221kPa 9:11
VENTING 9:11
AIR DRYING 10:11
CYCLE COMPLETE 25:11

Digital Signature #
FEB4C4DE14494E5A
