
Midmark M11 - v1.0.5
Total Cycles: 735

Name: RITTER2
Sterilizer ID: V2313228
01/31/2022 10:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.8 F 0.0
2:00 165.3 F 2.2
4:00 176.6 F 5.9
6:00 201.6 F 11.4
8:00 218.2 F 5.5
10:00 233.4 F 11.1
12:00 236.1 F 9.2
14:00 238.5 F 9.5
16:00 251.0 F 15.9
18:00 263.3 F 23.1
20:00 271.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.3 F 29.0
2:00 272.1 F 29.3
3:00 272.0 F 28.7
4:00 272.4 F 29.4

Min 271.4 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.5 F 0.0
5:00 231.2 F 0.0
10:00 194.7 F 0.0
15:00 188.0 F 0.0
20:00 185.2 F 0.0
25:00 209.2 F 0.0
30:00 225.1 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 20:03
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:58:28
CYCLE COMPLETE

01/31/2022 11:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____