

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3023

Name: RITTER 1  
Sterilizer ID: V2178198  
03/22/2023 09:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.8 F 0.0  
2:00 194.4 F 2.4  
4:00 193.7 F 6.6  
6:00 219.6 F 12.0  
8:00 227.7 F 7.8  
10:00 240.2 F 13.0  
12:00 240.3 F 11.1  
14:00 238.4 F 9.5  
16:00 252.8 F 16.7  
18:00 264.9 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.2 F 28.7  
2:00 271.9 F 28.8  
3:00 272.0 F 28.5  
4:00 271.6 F 28.7  
  
Min 271.2 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.4 F 0.0  
5:00 227.6 F 0.0  
10:00 243.8 F 0.0  
15:00 243.6 F 0.0  
20:00 234.3 F 0.0  
25:00 226.5 F 0.0  
30:00 224.1 F 0.0  
Drying Complete

FILLING: 1:38  
HEATING: 19:45  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:34  
CYCLE COMPLETE

03/22/2023 09:59 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_