
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001623
11:46 27/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:38
STERILIZING 3:36
136.0 C 220kPa 3:36
136.4 C 226kPa 4:06
136.4 C 225kPa 4:36
136.3 C 226kPa 5:06
136.2 C 224kPa 5:36
136.4 C 226kPa 6:06
136.4 C 225kPa 6:36
136.2 C 223kPa 7:06
136.2 C 223kPa 7:06
VENTING 7:06
AIR DRYING 7:57
CYCLE COMPLETE 8:03

Digital Signature #
3779D2320074B0EA
