

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000367  
8:00 21/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:40  
PRESSURIZING 2:46  
STERILIZING 3:55  
136.0 C 229kPa 3:55  
136.0 C 230kPa 4:25  
136.4 C 233kPa 4:55  
136.2 C 231kPa 5:25  
136.2 C 230kPa 5:55  
136.3 C 232kPa 6:25  
136.4 C 230kPa 6:55  
136.4 C 231kPa 7:25  
136.4 C 231kPa 7:25  
VENTING 7:25  
AIR DRYING 8:16  
CYCLE COMPLETE 8:35  
-----

Digital Signature #  
5804AEFC2B89B4CB  
-----