
Midmark M11 - v1.0.0
Total Cycles: 2681

Name: RITTER 1
Sterilizer ID: V2178198
12/15/2022 09:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.0 F 0.0
2:00 146.2 F 0.0
4:00 166.4 F 1.5
6:00 195.8 F 5.8
8:00 206.7 F 10.6
10:00 221.7 F 5.3
12:00 234.0 F 9.6
14:00 237.5 F 9.1
16:00 242.7 F 12.2
18:00 240.8 F 10.6
20:00 254.3 F 17.4
22:00 265.9 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.1
1:00 272.1 F 28.9
2:00 272.2 F 28.4
3:00 271.5 F 28.5
4:00 272.7 F 28.7

Min 271.0 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.8 F 0.0
5:00 234.9 F 0.0
10:00 217.4 F 0.0
15:00 232.0 F 0.0
20:00 223.3 F 0.0
25:00 210.4 F 0.0
30:00 236.3 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:30
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:38
CYCLE COMPLETE

12/15/2022 10:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____