
Midmark M11 - v1.0.0
Total Cycles: 4648

Name: RITTER 1
Sterilizer ID: V2178198
06/17/2024 11:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.2 F 0.0
2:00 174.2 F 2.3
4:00 191.7 F 6.0
6:00 204.2 F 10.6
8:00 217.1 F 5.4
10:00 226.0 F 8.2
12:00 237.0 F 12.7
14:00 236.3 F 9.5
16:00 239.7 F 9.9
18:00 246.7 F 13.6
20:00 258.1 F 19.8
22:00 268.4 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 29.0
2:00 271.5 F 28.3
3:00 272.5 F 29.2
4:00 272.1 F 28.7

Min 271.4 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 198.9 F 0.0
10:00 205.2 F 0.0
15:00 208.3 F 0.0
20:00 208.7 F 0.0
25:00 208.8 F 0.0
30:00 207.8 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 23:17
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:02:37
CYCLE COMPLETE

06/17/2024 12:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____