
Midmark M11 - v1.0.0
Total Cycles: 3084

Name: RITTER 1
Sterilizer ID: V2178198
04/10/2023 02:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 211.7 F 0.0
2:00 201.6 F 4.1
4:00 218.4 F 9.4
6:00 228.2 F 6.8
8:00 236.9 F 11.1
10:00 235.7 F 8.5
12:00 240.5 F 10.2
14:00 247.7 F 13.8
16:00 260.0 F 20.7
18:00 271.2 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.2
1:00 272.3 F 28.6
2:00 272.0 F 28.7
3:00 272.0 F 28.5
4:00 272.0 F 28.8

Min 271.3 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.6 F 0.0
5:00 235.7 F 0.0
10:00 231.5 F 0.0
15:00 228.5 F 0.0
20:00 228.6 F 0.0
25:00 227.5 F 0.0
30:00 230.5 F 0.0

Drying Complete

FILLING: 2:22
HEATING: 18:39
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:11
CYCLE COMPLETE

04/10/2023 03:48 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____