
Midmark M11 - v1.0.5
Total Cycles: 1470

Name: RITTER2
Sterilizer ID: V2313228
09/09/2022 08:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.7 F 0.0
2:00 74.2 F 0.0
4:00 86.2 F 0.0
6:00 111.4 F 0.0
8:00 147.5 F 0.0
10:00 184.4 F 3.3
12:00 203.4 F 7.4
14:00 218.0 F 11.8
16:00 224.4 F 6.6
18:00 234.8 F 11.0
20:00 234.6 F 7.8
22:00 243.5 F 11.9
24:00 245.5 F 13.1
26:00 257.9 F 19.9
28:00 269.0 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 271.8 F 28.9
2:00 272.0 F 29.1
3:00 272.2 F 29.2
4:00 272.2 F 29.2

Min 271.7 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 234.0 F 0.0
5:00 218.1 F 0.0
Dry Cycle Aborted

FILLING: 1:03
HEATING: 28:58
STERILIZING: 4:00
VENTING: 2:40
DRYING: 5:06
TOTAL CYCLE: 00:41:48

09/09/2022 08:58 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____