
Midmark M11 - v1.0.5
Total Cycles: 745

Name: RITTER2
Sterilizer ID: V2313228
02/01/2022 12:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.7 F 0.0
2:00 180.0 F 3.7
4:00 200.6 F 9.1
6:00 218.2 F 7.4
8:00 232.6 F 10.4
10:00 233.2 F 8.1
12:00 240.1 F 10.4
14:00 249.0 F 14.9
16:00 262.1 F 22.3
18:00 272.8 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.8 F 29.3
2:00 271.7 F 28.5
3:00 272.5 F 29.4
4:00 272.6 F 29.1

Min 271.5 F 28.4
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.7 F 0.0
5:00 219.9 F 0.0
10:00 215.8 F 0.0
15:00 214.9 F 0.0
20:00 211.9 F 0.0
25:00 208.6 F 0.0
30:00 205.5 F 0.0

Drying Complete

FILLING: 1:58
HEATING: 18:20
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 00:57:16
CYCLE COMPLETE

02/01/2022 01:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____