
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001747
9:13 14/06/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:46
STERILIZING 4:00
136.4 C 220kPa 4:00
136.3 C 224kPa 4:30
136.4 C 224kPa 5:00
136.0 C 222kPa 5:30
135.9 C 220kPa 6:00
136.2 C 223kPa 6:30
136.4 C 224kPa 7:00
136.3 C 223kPa 7:30
136.3 C 223kPa 7:30
VENTING 7:30
AIR DRYING 8:23
CYCLE COMPLETE 8:24

Digital Signature #
BDC04C891CEB709E
