

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1779

Name: RITTER 1  
Sterilizer ID: V2178198  
03/31/2022 02:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.4 F 0.0  
2:00 204.2 F 3.9  
4:00 209.3 F 9.2  
6:00 222.6 F 7.0  
8:00 233.2 F 10.7  
10:00 235.2 F 8.3  
12:00 239.8 F 10.1  
14:00 247.8 F 14.6  
16:00 260.9 F 21.9  
18:00 272.3 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.7 F 29.6  
2:00 272.0 F 28.9  
3:00 271.8 F 29.1  
4:00 272.8 F 29.7  
  
Min 271.5 F 28.7  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.6 F 0.0  
5:00 228.3 F 0.0  
10:00 233.7 F 0.0  
15:00 229.6 F 0.0  
20:00 228.0 F 0.0  
25:00 228.6 F 0.0  
30:00 228.3 F 0.0  
Drying Complete

FILLING: 2:24  
HEATING: 18:30  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:05  
CYCLE COMPLETE

03/31/2022 03:08 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_