

-----  
Midmark M11 - v1.0.5  
Total Cycles: 57

Name: RITTER2  
Sterilizer ID: V2313228  
05/31/2021 11:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.8 F 0.0  
2:00 188.7 F 3.2  
4:00 201.1 F 8.5  
6:00 217.8 F 9.7  
8:00 229.2 F 9.4  
10:00 235.8 F 9.9  
12:00 243.2 F 13.2  
14:00 244.0 F 12.6  
16:00 256.7 F 19.3  
18:00 267.7 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 272.3 F 29.6  
2:00 271.8 F 28.9  
3:00 271.4 F 29.1  
4:00 272.4 F 29.6  
  
Min 271.2 F 28.7  
Max 272.4 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.3 F 0.0  
5:00 240.3 F 0.0  
10:00 226.6 F 0.0  
15:00 221.9 F 0.0  
20:00 215.2 F 0.0  
25:00 213.1 F 0.0  
30:00 211.5 F 0.0  
Drying Complete

FILLING: 2:22  
HEATING: 19:24  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:60  
CYCLE COMPLETE

05/31/2021 12:21 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01