
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001482
16:14 25/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:33
STERILIZING 3:36
136.0 C 223kPa 3:36
136.3 C 225kPa 4:06
136.1 C 224kPa 4:36
136.3 C 226kPa 5:06
136.3 C 224kPa 5:36
136.4 C 227kPa 6:06
136.4 C 226kPa 6:36
136.4 C 226kPa 7:06
136.4 C 226kPa 7:06
VENTING 7:06
AIR DRYING 8:01
CYCLE COMPLETE 8:50

Digital Signature #
56ADB30731E34631
