
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 001590
14:13 11/17/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:49
PRESSURIZING 6:32
STERILIZING 8:05
136.0 C 220kPa 8:05
136.1 C 220kPa 8:35
136.4 C 223kPa 9:05
136.4 C 218kPa 9:35
136.1 C 220kPa 10:05
136.4 C 221kPa 10:35
136.4 C 222kPa 11:06
136.3 C 223kPa 11:35
VENTING 11:35
AIR DRYING 12:41
CYCLE COMPLETE 14:02

Digital Signature #
1109FF8ABE0C3C6F
