
Midmark M11 - v1.0.0
Total Cycles: 3472

Name: RITTER 1
Sterilizer ID: V2178198
06/30/2023 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.1 F 0.0
2:00 177.4 F 3.2
4:00 189.6 F 7.6
6:00 209.9 F 12.5
8:00 217.5 F 6.5
10:00 231.8 F 11.6
12:00 233.8 F 8.3
14:00 241.5 F 11.2
16:00 245.1 F 12.8
18:00 257.2 F 19.1
20:00 267.9 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 271.6 F 28.7
2:00 272.7 F 29.0
3:00 271.7 F 28.4
4:00 272.1 F 29.1

Min 271.0 F 28.1
Max 273.1 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 213.3 F 0.0
10:00 226.6 F 0.0
15:00 230.5 F 0.0
20:00 226.9 F 0.0
25:00 224.8 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 21:16
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 00:59:49
CYCLE COMPLETE

06/30/2023 12:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____