
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003529
11:49 26/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:23
PRESSURIZING 1:25
STERILIZING 2:32
136.0 C 220kPa 2:32
136.3 C 219kPa 3:02
136.4 C 224kPa 3:32
136.3 C 223kPa 4:02
136.4 C 224kPa 4:32
136.1 C 222kPa 5:02
136.0 C 221kPa 5:32
136.4 C 224kPa 6:02
136.4 C 224kPa 6:02
VENTING 6:02
AIR DRYING 7:01
CYCLE COMPLETE 7:03

Digital Signature #
59528188305E51DE
