

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4243

Name: RITTER 1  
Sterilizer ID: V2178198  
01/29/2024 03:48 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.5 F 0.0  
2:00 181.3 F 3.5  
4:00 195.6 F 8.3  
6:00 215.5 F 6.6  
8:00 232.7 F 9.8  
10:00 235.8 F 8.1  
12:00 245.6 F 12.8  
14:00 246.0 F 13.0  
16:00 259.3 F 20.1  
18:00 270.7 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.0  
1:00 271.8 F 28.3  
2:00 272.2 F 28.6  
3:00 272.4 F 28.8  
4:00 272.3 F 28.7  
  
Min 271.6 F 27.9  
Max 272.8 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 236.7 F 0.0  
Dry Cycle Aborted

FILLING: 1:43  
HEATING: 18:37  
STERILIZING: 4:00  
VENTING: 2:58  
DRYING: 0:19  
TOTAL CYCLE: 00:27:37

01/29/2024 04:15 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_