

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1417

Name: RITTER 1  
Sterilizer ID: V2178198  
12/01/2021 02:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.1 F 0.0  
2:00 158.2 F 2.5  
4:00 180.1 F 6.1  
6:00 199.9 F 10.9  
8:00 221.0 F 5.0  
10:00 235.2 F 10.3  
12:00 234.1 F 7.8  
14:00 244.1 F 12.8  
16:00 242.0 F 11.3  
18:00 255.2 F 18.0  
20:00 266.8 F 24.9  
22:00 272.9 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.7 F 29.1  
2:00 271.7 F 28.4  
3:00 272.4 F 29.1  
4:00 272.5 F 28.8

Min 271.3 F 28.3  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.0 F 0.0  
5:00 239.4 F 0.0  
10:00 222.2 F 0.0  
15:00 225.5 F 0.0  
20:00 225.4 F 0.0  
25:00 225.8 F 0.0  
30:00 223.6 F 0.0  
Drying Complete

FILLING: 1:48  
HEATING: 22:19  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:01:19  
CYCLE COMPLETE

12/01/2021 03:53 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02