
Midmark M11 - v1.0.5
Total Cycles: 2490

Name: RITTER2
Sterilizer ID: V2313228
06/08/2023 04:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.4 F 0.0
2:00 157.2 F 3.1
4:00 184.3 F 7.6
6:00 205.7 F 12.7
8:00 220.4 F 6.1
10:00 236.2 F 11.5
12:00 235.4 F 9.5
14:00 237.1 F 9.2
16:00 249.8 F 15.7
18:00 262.2 F 22.8
20:00 272.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.7 F 29.7
2:00 271.4 F 28.9
3:00 272.7 F 29.9
4:00 272.1 F 29.3

Min 271.3 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.7 F 0.0
5:00 216.1 F 0.0
10:00 220.8 F 0.0
15:00 226.6 F 0.0
20:00 232.6 F 0.0
25:00 222.0 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 20:18
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:07
CYCLE COMPLETE

06/08/2023 05:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____