
Midmark M11 - v1.0.5
Total Cycles: 4006

Name: RITTER2
Sterilizer ID: V2313228
10/07/2024 12:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.9 F 0.0
2:00 144.9 F 1.6
4:00 149.4 F 4.1
6:00 174.8 F 7.6
8:00 192.8 F 11.5
10:00 214.5 F 4.6
12:00 221.3 F 8.6
14:00 231.6 F 12.8
16:00 233.0 F 9.4
18:00 239.4 F 10.6
20:00 244.1 F 13.2
22:00 255.6 F 19.1
24:00 265.5 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 272.6 F 30.3
2:00 272.1 F 29.8
3:00 271.2 F 29.3
4:00 272.2 F 30.5

Min 271.0 F 29.3
Max 273.4 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 184.2 F 0.0
10:00 196.7 F 0.0
15:00 192.6 F 0.0
20:00 192.1 F 0.0
25:00 188.4 F 0.0
30:00 184.6 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:56
STERILIZING: 4:00
VENTING: 3:58
DRYING: 30:00
TOTAL CYCLE: 01:04:51
CYCLE COMPLETE

10/07/2024 01:17 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____