
Midmark M11 - v1.0.5
Total Cycles: 1115

Name: RITTER2
Sterilizer ID: V2313228
05/25/2022 12:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.6 F 0.0
2:00 130.3 F 0.4
4:00 155.5 F 3.0
6:00 177.6 F 6.7
8:00 203.6 F 11.1
10:00 214.8 F 5.0
12:00 228.4 F 9.8
14:00 234.2 F 9.1
16:00 238.8 F 11.5
18:00 236.8 F 9.4
20:00 249.7 F 15.6
22:00 260.5 F 21.8
24:00 270.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.7
1:00 272.0 F 29.8
2:00 271.8 F 29.8
3:00 272.0 F 29.7
4:00 272.4 F 29.7

Min 271.8 F 29.6
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.3 F 0.0
5:00 208.8 F 0.0
10:00 216.6 F 0.0
15:00 219.9 F 0.0
20:00 219.4 F 0.0
25:00 224.7 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:30
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:03:32
CYCLE COMPLETE

05/25/2022 01:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____