
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008420
7:48 19/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:30
STERILIZING 3:40
136.5 C 221kPa 3:40
136.3 C 223kPa 4:10
136.2 C 222kPa 4:40
136.3 C 218kPa 5:10
136.4 C 222kPa 5:40
136.4 C 220kPa 6:10
136.2 C 221kPa 6:40
136.4 C 222kPa 7:10
136.4 C 222kPa 7:10
VENTING 7:10
AIR DRYING 8:02
CYCLE COMPLETE 32:12

Digital Signature #
078DDBD5A0DC11CC
